



Contribution ID: 9

Type: **Talk**

The Coincident Doppler-Broadening Spectrometer at NEPOMUC

Tuesday 10 December 2019 15:00 (15 minutes)

The capabilities of the Coincident Doppler-Broadening Spectrometer at the Neutron-induced Positron Source Munich (NEPOMUC) will be presented. In addition an overview of planned expansions will be given.

Two beam modes are available: standard operations provide a beam spot of $> 300 \mu\text{m}$ (FWHM) and micro beam for scanning with a reduced diameter of $33 \mu\text{m}$ (FWHM). Measurements may either be conducted as standard Doppler-broadening spectroscopy (DBS) where the signal at each detector is treated separately or in coincident Doppler-broadening mode (CDBS) where the detectors are run as coincidence pairs. CDBS provides a greatly improved signal to noise ratio (10^8) allowing the analysis of the chemical surroundings of defect. Currently four different sample holders are available: i) a piezo x-y stage for precision 2D scanning and hence 3D defect imaging, ii) a heatable sample holder with $T_{max} = 1100 \text{ K}$ for T dependent defect spectroscopy, iii) a cryostat with $T_{min} = 40 \text{ K}$, iv) a device for in situ tensile tests. Ongoing improvement works include: an automated beam optimization system from which especially the micro beam mode will greatly profit. A planned increase in the number of detectors combined with an upgrade of the readout electronics, which should considerably reduce measurement times.

Authors: BURWITZ, Vassily Vadimovitch; CHRYSSOS, Leon; MATHES, Lucian

Co-authors: HUGENSCHMIDT, Christoph; SCHMIDT, Thomas

Presenters: BURWITZ, Vassily Vadimovitch; CHRYSSOS, Leon; MATHES, Lucian

Session Classification: Positrons

Track Classification: Positrons