

Contribution ID: 175

Type: Poster

The SoNDe high-flux neutron detector

Wednesday 9 December 2020 17:40 (20 minutes)

New high-flux and high-brilliance neutron sources demand a higher count-rate capability in neutron detectors. In order to achieve that goal, the Solid-State Neutron Detector (SoNDe) project developed a scintillation-based neutron detector. It is capable of fully exploiting the available flux current and coming neutron facilities, such as the European Spallation Source (ESS). [1] In addition to enabling high count-rates, one of the design goals was to develop a modular and scalable solution that can also be used in other instruments or different contexts, such as for laboratory setups. [2]

Since higher brilliance and flux sources call for detectors that can handle high-flux, especially when considering pulsed sources with high peak-flux, SoNDe provides

- Possible to handle a flux of more than 50 MHz on a 1x1 m^2 detector area
- Pixel resolution down to 3x3 mm^2
- Neutron detection efficiency higher than 80%, good gamma-discrimination
- μ s time resolution Count rates of 250 kHz per module (5 cm x 5 cm) were measured under primary beam conditions at neutron scattering experiments. Combined with the high area coverage of the square modules and the high efficiency of the scintillator this allows to use high flux neutron sources to capacity.

[1] JAKSCH, S., et al. Proceedings of the International Conference on Neutron Optics (NOP2017). 2018. S. 011019.

[2] JAKSCH, S., et al. Cumulative Reports of the SoNDe Project July 2017. arXiv preprint arXiv:1707.08679, 2017.

Authors: JAKSCH, Sebastian (Physicist); FRIELINGHAUS, Henrich (JCNS); ENGELS, Ralf; DESERT, Sylvain (CEA Saclay - LLB); Prof. HALL-WILTON, Richard (European Spallation Source ERIC (ESS)); Ms PERREY, Hanno (Lund University); GHEORGHE, Codin (Integrated Detector Electronics AS)

Presenter: JAKSCH, Sebastian (Physicist)

Session Classification: Joint poster session of MLZ User Meeting and DN2020

Track Classification: DN: Instrumentation