German Conference for Research with Synchrotron Radiation, Neutrons and Ion Beams at Large Facilities



Contribution ID: 358

Type: Talk

Dose efficient, high resolution X-ray imaging applied to biological model organisms

Monday, 17 September 2018 11:45 (15 minutes)

The main motivation of our work is to develop high-throughput as well as dose efficient X-ray imaging instrumentation and techniques for small animal imaging of vertebrate model organisms with high resolution and adjustable field of view (FOV) for multi-scale observations of whole organisms, organs, and cellular processes. High-resolution 3D and 4D X-ray imaging of model organisms and their developmental processes provides valuable insights and important information for life sciences without the need for dissecting the specimen, thus also allowing for *in vivo* studies. However, especially in the case of synchrotron *in vivo* experiments, the dose impinging on the specimen is crucial and has to be minimized.

Methodical routes for dose-efficient *in vivo* studies enclose in-line phase contrast imaging (PCI) and the use of Bragg Magnifier (BM) optics. By using asymmetric Bragg reflection, we achieve a magnification of up to 200 in 2D. By placing the BM downstream of the sample and combining it with a photon counting pixel detector, we obtain a highly resolving (sub- μ m) and very dose efficient X-ray microscope. By placing the BM upstream of the sample, we can adjust the FOV up to several cm² and preserve or even enhance the coherence properties of synchrotron beamlines.

In this contribution we show exemplary results of high-throughput imaging of whole Medaka, *in vivo* PCI measurements of *Xenopus* embryos, as well as the design and first experimental results of the BM instrumentation.

Primary authors: Dr HAMANN, Elias (Institute for Photon Science and Synchrotron Radiation (IPS), Karlsruhe Institute of Technology (KIT)); Dr BELLUCCI, Valerio (Laboratory for Application of Synchrotron Radiation (LAS), Karlsruhe Institute of Technology); Mr HESSDORFER, Holger (Institute for Photon Science and Synchrotron Radiation (IPS), Karlsruhe Institute of Technology (KIT)); Ms ENGELHARDT, Sabine (Centre for Organismal Studies (COS) Heidelberg; Institute for Photon Science and Synchrotron Radiation (IPS), Karlsruhe Institute of Technology (KIT)); Mr ZUBER, Marcus (Laboratory of Applications of Synchrotron Radiation (LAS), Karlsruhe Institute of Technology); Dr HOFMANN, Ralf (Institute for Photon Science and Synchrotron Radiation (IPS), Karlsruhe Institute of Technology (KIT)); Dr VAGOVIC, Patrik (DESY, XFEL); Mr HRIVŇAK, Stanislav (Department of Biophysics, Faculty of Science, P. J. Šafárik University, Košice); Dr THUMBERGER, Thomas (Centre for Organismal Studies (COS) Heidelberg); Prof. WITTBRODT, Joachim (Centre for Organismal Studies (COS) Heidelberg); Prof. BAUMBACH, Tilo (Karlsruhe Institute of Technology)

Presenter: Dr HAMANN, Elias (Institute for Photon Science and Synchrotron Radiation (IPS), Karlsruhe Institute of Technology (KIT))

Session Classification: Parallel session 1

Track Classification: P1 Instrumentation and methods