



Contribution ID: 327

Type: **Poster**

## Larmor diffraction -status and new concepts

*Tuesday, 18 September 2018 17:15 (15 minutes)*

We review the current status of Larmor diffraction (LD) at TRISP at the MLZ. The current setup with precession fields both up- and downstream the sample has a good resolution and is not sensitive to the crystal size or beam divergence. The disadvantage is that it is not possible to work with spin-depolarizing ferromagnetic samples or with magnetic fields at the sample. There are two proposals for a LD configuration using precession fields only along the incident beam upstream the sample, such that spin-depolarization at the sample is not harmful. The first of these new concepts is the single-arm LD proposed by Rekveldt and van Well, the second is a spin modulation technique using one radio-frequency spin-flipper (Habicht, unpublished). Both new concepts loose resolution in comparison to the classical LD. We will discuss and compare these new LD methods.

**Primary authors:** KELLER, Thomas (MPI for Solid State Research); HABICHT, Klaus (Helmholtz-Zentrum Berlin); KEIMER, Bernhard (MPI for Solid State Research)

**Presenter:** KELLER, Thomas (MPI for Solid State Research)

**Session Classification:** Poster session 2

**Track Classification:** P1 Instrumentation and methods